Public Health Service Food and Drug Administration

91384d

19900 MacArthur Blvd., Ste 300 Irvine, California 92612-2445 Telephone (949) 798-7600

WARNING LETTER

Certified Mail Return Receipt Requested

June 14, 2001

Carolyn Rhee
Associate Administrator
LAC/ Olive View Medical Center
Department of Radiology
14445 Olive View Drive
Sylmar, CA 91342

W/L Number: 58 - 01

Inspection ID: 1290640010 CFN: 20-29,626

FEI: 1000518951

Dear Ms Rhee:

We are writing to you because on June 4, 2001 your facility was inspected by a representative of the State of California acting in behalf of the U. S. Food and Drug Administration (FDA). This inspection revealed a serious regulatory problem involving the mammography at your facility.

Under a United States Federal law, the Mammography Quality Standards Act of 1992 (MQSA), your facility must meet specific requirements for mammography. These requirements help protect the health of women by assuring that a facility can perform quality mammography. The inspection revealed the following Level 1 findings at your facility:

- Level 1: Phantom quality control (QC) records were missing for at least 4 weeks for unit #1 (an machine, model machine, model machine, serial number which is located in room A.
- Level 1: Phantom QC records were missing for at least 4 weeks for unit #2 (a machine, model number serial number which is located in room B.

The specific problems noted above appeared on your MQSA Facility Inspection Report, which was issued to your facility at the close of the inspection. These problems are identified as Level 1 because they identify a failure to meet a significant MQSA requirement.

Page Two of Three June 14, 2001

re: LAC/ Olive View Medical Center re: Warning Letter Number 58 - 01

Because these conditions may be symptomatic of serious underlying problems that could compromise the quality of mammography at your facility, they represent a serious violation of the law which may result in FDA taking regulatory action without further notice to you. These actions include, but are not limited to, placing your facility under a Directed Plan of Correction, charging your facility for the cost of on-site monitoring, assessing civil money penalties up to \$10,000 for each failure to substantially comply with, or each day of failure to substantially comply with, MQSA Standards, suspension or revocation of your facility's FDA certificate, or obtaining a court injunction against further mammography.

In addition, your response should address the Level 2 findings that were listed on the inspection report provided to you at the close of the inspection. These Level 2 findings are:

- Level 2: Failed to produce documents verifying that the interpreting physician, met the continuing experience requirement of having interpreted or multi-read 960 mammograms in twenty-four (24) months.

It is necessary for you to act on this matter immediately. Please explain to this office, in writing, within fifteen (15) working days from the date you received this letter:

- the specific steps you have taken to correct all of the violations noted in this letter;
- each step your facility is taking to prevent the recurrence of similar violations; and
- please provide sample records that demonstrate proper record keeping procedures, if the findings relate to quality control or other records (Note: Patient names or identification should be deleted from any copies submitted).

Please submit your response to:

Thomas L. Sawyer
Director, Compliance Branch
U.S. Food & Drug Administration
19900 MacArthur Blvd.; suite #300
Irvine, CA 92612-2445
Phone: (949) 798-7600

Finally, you should understand that there are many FDA requirements pertaining to mammography. This letter pertains only to findings of your inspection and does not necessarily address other obligations you have under the law. You may obtain general

Page Three of Three June 14, 2001

re: LAC/ Olive View Medical Center re: Warning Letter Number 58 - 01

information about all of FDA's requirements for mammography facilities by contacting the Mammography Quality Assurance Program, Food and Drug Administration, P.O. Box 6057, Columbia, MD 21045-6057 (telephone number 1-800-838-7715) or through the Internet at http://www.fda.gov.

If you have more specific questions about mammography facility requirements, or about the content of this letter, please feel free to contact Beverly Thomas (MQSA Auditor) at telephone number 949-798-7708.

Sincerely,

Alonza E. Cruse District Director

cc:

State of California
Dept. of Health Services
Radiological Health Unit
550 South Vermont Avenue; Suite #601
Los Angeles, CA 90020